

Test Report issued under the responsibility of :



TEST REPORT	
United Nations, Recommendations on the Transport of Dangerous Goods, Manual of Test and Criteria (Rev. 5th, Amendment 1), Section 38.3	
Report Reference No	128689-658487B
Total number of pages	25
Testing Laboratory	Laboratoire Central des Industries Electriques (LCIE)
Address.....	33 Avenue du General Leclerc – 92260 FONTENAY-AUX-ROSES (FRANCE)
Applicant's name	SCHILLER MEDICAL
Address	4 rue Louis Pasteur 67162 WISSEMBOURG - France
Test specification:	
Standard.....	United Nations, Recommendations on the Transport of Dangerous Goods, Manual of Test and Criteria (Rev. 5th, Amendment 1), Section 38.3.
Test procedure	--
Non-standard test method.....	N/A
Test item description	Cylindrical Lithium primary cell
Trade Mark	DURACELL
Manufacturer	Procter & Gamble Switzerland SARL (DURACELL)
Model/Type reference	123A (CR17345)
Ratings	3Vdc - 1500 mAh - Li/MnO2 cell

Date of issue

Fontenay-aux-Roses, on 2014-10-27

Conclusion : The sample complies with the requirements of the tests performed.

Technical manager

Alain POSLUSZNY

This document shall not be reproduced, except in full, without the written approval of the LCIE. This document contains results related only to the item tested. It does not imply the conformity of the whole production to the item tested. Unless otherwise specified, the decision of conformity takes into account the uncertainty of measures. This document does not anticipate any certification decision.

LCIE
Laboratoire Central
des Industries Electriques
Une société de Bureau Veritas

33, avenue du Général Leclerc
B.P. n° 8
92266 Fontenay-aux-Roses cedex
France

Tél : +33 1 40 95 60 60
Fax : +33 1 40 95 86 56
contact@lcie.fr
www.lcie.fr

Société par Actions Simplifiée
Au capital de 15 745 984 €
RCS Nanterre B 408 363 174

Summary of testing :																	
<p>Tests performed (name of test and test clause) :</p> <table border="1"> <thead> <tr> <th>Clause</th> <th>Contents of Test</th> </tr> </thead> <tbody> <tr> <td>38.3.4.1</td> <td>Test T.1: Altitude simulation</td> </tr> <tr> <td>38.3.4.2</td> <td>Test T.2: Thermal test</td> </tr> <tr> <td>38.3.4.3</td> <td>Test T.3: Vibration</td> </tr> <tr> <td>38.3.4.4</td> <td>Test T.4: Shock</td> </tr> <tr> <td>38.3.4.5</td> <td>Test T.5: External short circuit</td> </tr> <tr> <td>38.3.4.6</td> <td>Test T.6: Impact</td> </tr> <tr> <td>38.3.4.8</td> <td>Test T.8: Forced discharge</td> </tr> </tbody> </table>	Clause	Contents of Test	38.3.4.1	Test T.1: Altitude simulation	38.3.4.2	Test T.2: Thermal test	38.3.4.3	Test T.3: Vibration	38.3.4.4	Test T.4: Shock	38.3.4.5	Test T.5: External short circuit	38.3.4.6	Test T.6: Impact	38.3.4.8	Test T.8: Forced discharge	<p>Testing location :</p> <p>Laboratoire Central des Industries Electriques (LCIE) 33 Avenue du General Leclerc 92260 FONTENAY-AUX-ROSES (FRANCE)</p>
Clause	Contents of Test																
38.3.4.1	Test T.1: Altitude simulation																
38.3.4.2	Test T.2: Thermal test																
38.3.4.3	Test T.3: Vibration																
38.3.4.4	Test T.4: Shock																
38.3.4.5	Test T.5: External short circuit																
38.3.4.6	Test T.6: Impact																
38.3.4.8	Test T.8: Forced discharge																
<p>Summary of compliance with National Differences:</p> <p>None</p>																	